Question n. 1: In Lot1. Medical Equip. Dental unit (There is no technical specs mentioned in the file).

Answer n. 1: Please refer to the following specifications

Item number	Item Name	Specifications required	Specifications Offered	Notes, remarks,
				ref to documentation
02-DEU	DENTAL UNIT	Dental Chair with Compressor.		
		(Oil-less, 24ltr air tank capacity ca, robust construction.		
		Electric driven chair (Preferable)		
		High density foam rubber upholstery		
		Auto zero position with at least four programmable positions		
		Up/down backrest		
		Built in safety switches		
		Multi function foot-switch		
		Low heat halogen light with removable handle for sterilization		
		Two dental stools, 5 castors, revolvable, height adjustable with backrest		
		The Dentist element is to have fibre optic light built in the handsets, and to feature;		
		High speed turbine, $> 200,000$ rpm with quick coupling		
		Low speed micro-motor with two hand pieces, contra angle 1:1, straight 1:1		
		High power ultrasonic scalar, regulated		
		Three way water and air syringe		
		Curing light		
		Instrument tray		
		Built in x-ray light viewer		
		The assistant element is to feature:		
		Assistant control key switches		
		High volume suction		
		Low volume suction		
		Cup filler		
		Rinsing cuspidor		
		Built in automatic separator to be connected with a dry suction system		

Item number	Item Name	Specifications required	Specifications Offered	Notes, remarks,
number				
				ref to documentation
		Automatic disinfection unit		
		Power line : $220 \text{ V} \pm 20\%$, 50 Hz.		

QUESTION n.2: In Lot 1. Surgical Inox Set. Set, Ear Syringe (the item mentioned in technical specs. table but not mentioned the quantity table)

ANSWER n. 2: The Item is not required and has not to be quoted.

QUESTION n. 3 The location of centers (Alzihur, Teraat Albija, Shambat) in which City, in order to enable us calculate the domestic shipping cost.

ANSWER n. 3: The health center are all located in KHARTOUM