



CLARIFICATION n.2

CIG NUMBER: 8643078

CONTRACT NAME: Supply of Biomedical Equipment for three Health Centres in Shambat, al Zuhur e Teriat Al Bija in Khartoum in Sudan – SUDAN

SUBJECT: Clarification to Technical Specifications

Clarification n. 1

In Section B - Technical Specifications (*3HC c4f1_annexiitechspeciitechoffer_en*) are mentioned the technical Specifications of the following items:

- LOT 1 “**COUCH EXAMINATION, TREATMENT**”
- LOT 1” **SET, EAR SYRINGE**”

No quantity is required for those two items therefore the Tenderer will disregard the corresponding technical specifications.

Clarification n. 2

In Section B - Technical Specifications (*3HC c4f1_annexiitechspeciitechoffer_en*), where is mentioned the item in Lot 1:

- OPTHALMOSCOPE/RETINOSCOPE,

please read:

- RETINOSCOPE

Clarification n. 3

In Section B - Technical Specifications (*3HC c4f1_annexiitechspeciitechoffer_en*), the Item DRUM, STERILIZING INOX, **MEDIUM** is mentioned twice. In the second case please consider the following dimensions:

- size approximately (**25***36) cm

and not **small** size, [...] size approximately (**15***36) cm.



Clarification n. 4

In Section B - Technical Specifications (3HC c4f1_annexiitechspeciitechoffer_en), Lot1, consider the following amendment:

- TRAY, INSTRUMENT, INOX, LARGE: The indicative dimensions are: 30 cm Length; 22 cm Width; 0,5 Depth
- TRAY, INSTRUMENT, INOX, SMALL: The indicative dimensions are: 23 cm Length; 15 cm Width; 0,5 Depth

Clarification n. 5

In Section B - Technical Specifications (3HC c4f1_annexiitechspeciitechoffer_en), Lot1, for the item described as

- SET, SMALL FRAGMENT/FOREIGN BODIES: Various Size.

Please consider the following technical Specifications: *Forceps for extraction of small fragments from eye; 4 pieces various shape and size (indicative length from 7 cm to 15 cm).*

Clarification n. 6

Wherever it appears the Item Name 03-ThDL 100 6 600 must be read as **Thermometer, Data Logger, w/ display.**

Clarification n. 7

In Section B - Technical Specifications (3HC c4f1_annexiitechspeciitechoffer_en), Lot1, the item Dental Unit **is required** in the stated quantity indicated in the list at *SECTION B - Instructions to tenderers – Art. 1* and must be quoted considering the following technical Specifications, using the following format:

Item number	Item Name	Specifications required	Specifications Offered	Notes, remarks, ref to documentation
02-DEU	DENTAL UNIT	Dental Chair with Compressor. (Oil-less, 24ltr air tank capacity ca, robust construction. Electric driven chair (Preferable) High density foam rubber upholstery Auto zero position with at least four programmable positions Up/down backrest Built in safety switches Multi function foot-switch Low heat halogen light with removable handle for sterilization		



Item number	Item Name	Specifications required	Specifications Offered	Notes, remarks, ref to documentation
		<p>Two dental stools, 5 castors, rolling, height adjustable with backrest</p> <p>The Dentist element is to have fiber optic light built in the handsets, and to feature;</p> <p>High speed turbine, > 200,000 rpm with quick coupling</p> <p>Low speed micro-motor with two hand pieces, contra angle 1:1, straight 1:1</p> <p>High power ultrasonic scalar, regulated</p> <p>Three way water and air syringe</p> <p>Curing Light</p> <p>Instrument tray</p> <p>Built in x-ray light viewer</p> <p>The assistant element is to feature:</p> <p>Assistant control key switches</p> <p>High volume suction</p> <p>Low volume suction</p> <p>Cup filler</p> <p>Rinsing cuspidor</p> <p>Built in automatic separator to be connected with a dry suction system</p>		
		<p>Automatic disinfection unit</p> <p>Power line : 220 V \pm 20% , 50 Hz.</p>		