

Clarification n.3

REF: 8643078 CIG-931926224A Lot 3 : Radiology systems

CONTRACT NAME: Supply of Biomedical Equipment for three Health Centres in Shambat, al Zuhur e Teriat Al Bija in Khartoum in Sudan – SUDAN
SUBJECT: Clarification to Technical Specifications

Question n. 1

Item 02-RU-F : the generator requested is minor 125 KV which came with fixed table .The table requested is movable 4-way floating top type with foot switch control which for major generator 150KV. Please advise which specifications should be changed: Generator or table?

Reply n.1

The technical specifications and the requested configuration are confirmed.
The examination table and the X-ray tube Power are configurable elements of the system.

The deadline to submit the tender dossier is **18/09/2022 Khartoum Time: 17.00**